



PT PINDAD

NOMOR (Number) : AC 873 /PM/RM/II/2024  
 EDISI/REV. (Edition/Rev.) : 1 / 0  
 TANGGAL (Date) : 28 FEBRUARI 2024  
 HALAMAN (Page) : 1 DARI (of) 1

**KRITERIA KEBERTERIMAAN  
(ACCEPTANCE CRITERIA)**

1. NAMA PRODUK (Product Name) : CMM Portable
2. NOMOR GAMBAR (Drawing Number) : N/A (Produk Standar)
3. NOMOR MATERIAL (Material Number) : N/A (Produk Standar)

A. SPESIFIKASI (Specification) :



**PMN 2023**  
**PT PINDAD**

CONTROLLED D  
2024

Gambar ilustrasi CMM Portable setara

NO	SPECIFICATION	
1	Model	Romer Absolute Arm 8330- 7 axis
2	Measuring Range	3 m
3	Wrist display	Touch screen
4	Axis configuration	7 axis probing
5	Working temperature	+5 °C to 40°C
6	Storage temperature	-30 °C to 70°C
7	Operating humidity range	10 to 90% non condensing
8	Touch Probes supplied as standart	3 mm, 5 mm and 15 mm
9	Compatible with tube probe	6 tube probes are offered
10	Probe security (when arm is not use)	Homedock
11	Power supply	100 - 240 Vac
12	Protecting rating	IP54
13	Accuracy	According ISO 10360 - 12

B. PERSYARATAN DOKUMEN MUTU (*Quality Document Requirement*) :

1. Sertifikat of Origin
2. Sertifikat kalibrasi dari lembaga yang terakreditasi
3. Sertifikat / Surat Garansi selama 2 tahun
4. Surat pernyataan ketersediaan sukucadang selama 10 tahun

C. UJI KEBERTERIMAAN (*Acceptance Test*) :

NO (Nr)	MATA UJI (Item Test)	PERSYARATAN (Requirement)	ALAT UKUR (Equipment)	JUMLAH SAMPLING (Sampling Number)
1	Visual Produk	Kondisi baru, tidak cacat	Visual	Sensus
2	Kesesuaian spesifikasi	Sesuai poin A	Visual	Sensus
3	Uji Fungsi	Sesuai spesifikasi	Uji coba	sensus
4	Keseuaian Dokumen	Sesuai Poin B	Visual	Sensus

	DISIAPKAN ( <i>Prepared</i> )	DISETUJUI ( <i>Approved</i> )
JABATAN ( <i>Title</i> )	EXPERT BID ENG MUTU PRODUK INDUSTRIAL	MANAGER REKAYASA MUTU
N A M A ( <i>Name</i> )	YUDI YULIANSAH	S U P A R N O
TANGGAL ( <i>Date</i> )	28 FEBRUARI 2024	28 FEBRUARI 2024
TANDA TANGAN ( <i>Signature</i> )		

